

Corrigendum Pre Bid Clarifications

Tender No: RCB/NIT/18/20-21/IS

Tender Name: supply, installation and commissioning of Incubator Shaker

S. No.	Name of the Agency	Bidding Document Reference(s) (Section number/page number)	Content of RFP requiring clarification	Clarification raised by the bidder	RCB Remarks
1	Eppendorf	Technical Specification Point No. 2 General Specifications	Capacity of incubator	238 liters or above	Point accepted
2	Eppendorf	Technical Specification Point No..14 performance	Installation all over India	Quoted model must have at least 10 installations all over India,	Point accepted
3	Eppendorf	Technical Specification Point No..15 Training	On-site training	Company should have dedicated application specialists with a proven track record to provide onsite training at the time of installation.	No change
4	Eppendorf	Technical Specification Point No..16 b Warranty	Post warranty CMC (after 3 year)	The post- warranty CMC (after 3 years) for 5 year from OEM should also be quoted for spares and labor for the complete system, which includes all the peripherals and accessories.	Point accepted
5	Eppendorf	<u>Important Note:</u>	price comparison	The price comparison shall be made take into account the cost of the equipment and accessories with 3-year warranty and post warranty CMC for Five years. Failure to comply this condition entails rejection of bids.	Point accepted