

Addendum1

Tender No.: RCB/GTE/03/22-23/AFC

The competent authority has approved following further changes in the technical specifications:

Tender Clause Ref:	Existing	May be read as
Section VII	Pumping system with maximum flow rate of 300 ml/min or more & amp; maximum pressure capacity of 150 psi or more.	Pumping system with maximum flow rate of 250 ml/min or more & amp; maximum pressure capacity of 150 psi or more.

-SD XXX-

Tender Inviting Authority