

Corrigendum

The competent authority has approved following changes in the technical specifications:

Tender Clause Ref:	Existing	May be read as
Section VII Point No. 2(a)	UV-Vis detector should monitor at least three wavelengths in 200 – 750 nm range	UV-Vis detector should monitor at least three wavelengths in 200 – 700 nm or more range

-SD XXX-

Tender Inviting Authority