

Pre-bid meeting clarification of Plant Growth Room, held on 16.08.2023

Sr. No.	Tender Specification	Point raised by the Sreekara Envirotech Pvt Ltd Company	Approved/Justification
1	Relative humidity (RH) range at 21°C-35°C: 65%-85% RH with accuracy ±2% RH, sensed by electronic capacitive humidity sensor mounted in the room.	Relative (RH) range at 21°C-35°C: 65%-85% RH with accuracy ±3% RH, sensed by electronic capacitive humidity sensor mounted in the room.	The RH accuracy of ±3% may be approved.
2	Light intensity over each growth area must be at least 300 moles (20000 lux) measured with the light sensors positioned at the canopy above each shelf. Electronic ballasts with 8 Nos of cool white Fluorescent Lights should be present in each canopy on each tier. Programming and control of photoperiod (Light/Dark cycle) should be performed by the central Controller.	Light intensity over each growth area must be at least 300 moles (20000 lux) measured at a distance of 6 inches from the light source. The cool white Fluorescent light may be replaced with LED lights.	Accepted. The measured distance for light intensity should be at 6 inches below the light source from the canopy. And LED lights must be used instead of cool white fluorescent lights.
3	The Delivery Period is maximum 90 days from date of issue of Purchase Order against the Contract.	As this is a relatively large project with 4 Nos of Plant Growth chambers to be delivered, installed, commissioned and handed over, you are requested to reconsider the delivery period to 180 days from the date of receipt of the purchase order	Accepted, The Delivery Period may be extended to a maximum of 180 days from date of issue of the Purchase Order against the Contract.
4	100% payment would be made on receipt of goods in good condition and acceptance	Kindly confirm that 30% advance payment shall be released to the successful bidder upon order acceptance against	Payment terms may be read as 30% advance and balance payment shall be made after successful installation ,commissioning and acceptance of the project

		submission of proforma invoice.	
5	The bidders who are currently registered with MSME for the goods as per Tender document specification shall be eligible for exemption from Bid Security as defined in MSE Procurement Policy issued by the department of MSME. In case the bidder falls in this category, the bidder shall upload relevant certificate of registration for the subject goods issued by department of MSME	EMD and Processing Fees shall be exempted for MSME registered bidders	EMD is exempted for MSME/NSIC registered vendors. However exemption of processing fee is as per eWizard rules.
06	Statement of financial standing from C.A. or Bank with address & proof of turnover of the firm minimum 25 lacs for the last 3 years	Turnover 25 Lacs shall be cumulative amount of the last three financial years turnover.	Accepted, Considering many recent startups, we may have to relax this clause to encourage more participation in the tender from Indian companies.
Additional request		Pre requisites like complete civil works required for the equipment, power supply points up to the chamber, R.O unit & AC units, complete water line up to the R.O unit and Platforms for RAC will be in the scope of the purchaser.	Accepted.
Additional request		Electrical consumables like Lights & Ballasts and consumables like filters will not be covered under warranty.	Not accepted, All items installed should be covered during the warranty period.
Additional request		We will quote for CMC after warranty period as per the terms of the tender. However, you are requested not to	Not accepted, Price of CMC shall be taken into account for determining L1 bidder

		consider the CMC amount for L1 calculations	
--	--	---	--

Another representative during the pre-bid meeting was from **M/s Saveer Biotech Ltd.** The concerns and doubts raised by the representative (attached) was clarified during the pre-bid meeting itself. In their Point#2, the representative requested for LED lights instead of cool white fluorescent lights regulated by electronic ballasts. This suggestion is accepted and revised (Point #2 in the above table). There were no further changes requested.