

REGIONAL CENTRE FOR BIOTECHNOLOGY

Regional Centre for Biotechnology,
at NCR Biotech Science Cluster, Faridabad, Village-Bhankri,
Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon,
District Faridabad (Haryana)- 121001 Tele No 0129-2876109

(SCHEDULE-'A')

S.NO. OF TENDER : _____
FILE NO. : RCB/ATN/10/14-15
Name of the party in whose : _____
Favour the Tender form has : _____
been issued : _____

The Executive Director,
Regional Centre for Biotechnology,
at NCR Biotech Science Cluster, Faridabad, Village-Bhankri,
Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon, District
Faridabad (Haryana)- 121001 (Tele No 0129-2876109)

(SEAL OF THE OFFICER)

Dear Sir,

1. I/We hereby submit our tender for the _____
2. I/WE now enclosing herewith the Bank Guarantee/D.D. No..... dated..... for **Rs 1,40,000.00** in favour of the "EXECUTIVE DIRECTOR, RCB" towards EMD/Bid Security. (TENDERS NOT ACCOMPANIED WITH EMD/BID SECURITY ALONGWITH THE TECHNO-COMMERCIAL BID SHALL BE SUMMARILY REJECTED).
3. I/We have gone through all terms and conditions of the tender documents before submitting the same.
4. I/We hereby agree to all the terms and conditions, stipulated by the RCB, in this connection including delivery, warranty, penalty etc. Quotations for each group are being submitted under separate covers, and sheets and shall be considered on their face value.
5. I/We have noted that overwritten entries shall be deleted unless duly cut & re-written and initialed.
6. Tenders are duly signed (No thumb impression should be affixed).
7. I/We undertake to sign the contract/agreement, if required, within 15 (Fifteen days) from the date of issue of the letter of acceptance, failing which our/my security money deposited may be forfeited and our/my name may be removed from the list of suppliers at the RCBGurgaon.

NOTE: ALL TERMS & CONDITIONS SUCH AS TAXES ETC, HAS BEEN INDICATED IN THE QUOTATIONS FAILING WHICH IT WILL BE PRESUMED THAT THE RATES ARE INCLUSIVE OF ALL TAXES AND OTHER TERMS AND CONDITIONS ARE ALSO AS PER YOUR REQUIREMENTS.

Yours faithfully,

Signature of
Tenderer(s) full Address.

WITNESS _____
WITNESS _____

ADVERTISED TENDER NOTICE

**New Campus at Regional Centre for Biotechnology, at NCR Biotech Science Cluster,
Gurgaon-Faridabad Express Way, Gurgaon-Faridabad Expressway 3rd Milestone
towards Gurgaon, Village-Bhankri, District Faridabad (Haryana) 121001
Tele No 0129-2876109**

RCB/ATN/08-11/14-15

Dated 19.03.2015

NOTICE INVITING TENDER

1. Sealed Tenders are invited, under **TWO-BID SYSTEM**, from reputed manufacturers / suppliers for the supply and installation of following items at **New Campus at Regional Centre for Biotechnology, at NCR Biotech Science Cluster, Gurgaon-Faridabad Express Way, Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon, Village-Bhankri, District Faridabad (Haryana)**

<u>Tender Ref.</u>	<u>Item/Equipment</u>	<u>Qty</u>	<u>Tender Fee (in INR)</u>	<u>Bid Security/EMD (In INR)</u>
RCB/ATN/08/14-15	FPLC System	01 (One)	1000/-	Rs, 70,000/-
RCB/ATN/09/14-15	Microbial Cell Homogenizer	01(One)	1000/-	Rs 30,000/-
RCB/ATN/10/14-15	Bioreactor (7 Ltrs and 14 Ltrs)	02 (Two) one each	1000/-	Rs 1,40,000/-
RCB/ATN/11/14-15	Controlled WALK-IN-PLANT Growth Chamber	01(one)	1000/-	Rs 40,000/-

2. Copy of Tender document and other details may be downloaded from www.eprocure.gov.in (CPP Portal), www.rcb.res.in and www.rcb.ac.in. Sealed Bids (Two-Cover System) will be received upto 14:00 hrs on or before 21.04.2015 and opening of Tenders will be on same day (21.04.2015) at 15:00 hrs.

(SR. MANAGER)
For & on behalf of
EXECUTIVE DIRECTOR

CHECK LIST FOR TERMS AND CONDITIONS

Check list for Terms and Conditions (To be filled by the bidder and submitted alongwith the technical bid)

1. Certificate for being in business for more than 2 years _____
2. Certificate for sole ownership/partnership _____
3. Statement of financial standing from bankers _____
4. Performance report/List of organization supplied with the same equipment _____

5. Whether rates quoted included all taxes/Rate is a CIF _____
6. Whether rates are quoted as per tenders specifications _____
7. Authority letter from manufacturer/principal enclosed _____
8. Affidavit that the firm has not been black listed in the past by any Organization _____
9. Affidavit that the firm has no vigilance case/CBI/FEMA case pending against him/supplier(principal)

10. Affidavit that the firm is not supplying the same item at the lower rate than the rate quoted in the tender to any Govt. organization or any other Institute (Fall clause) _____
11. Quotation being submitted directly by the manufacturer or authorized distributor

12. Statement of turnover/annual report for each of the last three years are attached

13. Bid Security amount deposited is enclosed _____
14. Literature of original catalogue of the product is attached for reference _____
15. Details of Name of beneficiary, Account No. of the beneficiary, IFCS code of the bank/branch enclosed at Page No. _____
16. Comprehensive Guarantee/Warranty period for _____ and thereafter-
Comprehensive AMC (including all spares & labour) YES/NO. (Tick the option YES/NO) for
further years.
17. Compliance Statement with relation to specification _____.

**(NAME OF THE BIDDER)
WITH SIGNATURE & SEAL**

REGIONAL CENTRE FOR BIOTECHNOLOGY
NCR Biotech Science Cluster, Gurgaon-Faridabad Express Way,
Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon,
Village-Bhankri, District Faridabad (Haryana) Tele No 0129-2876109

F. No. RCB/ATN/10/14-15

Specifications for 2 independent in situ fermenters (7 L and 14 L capacity) for Molecular Interaction and Protein Purification Facility (ATPC)

I Fermentor with 7L total capacity

1. Fermentor:

- In situ sterilizable fermentor with working volume 2 to 5 litre (Total Volume 7 litre). Fermentor should be capable of accommodating 14 L and 19.5 L vessel
- System should be suitable for microbial fermentation in aerobic & anaerobic condition for batch, fed batch & continuous operation
- Aspect Ratio: 2:1
- The fermentor should have advanced synoptic touch screen interface providing easy to read displays of current values, cascade loops sterilization cycles, customized PI values for all process parameter etc.

2. Process Definition:

- Programmable in situ sterilization, the system should be supplied with Stainless Steel Exhaust Condenser on headplate.
- Fermentor should be capable of reaching sterilization temperature up to at least 121°C and temperature control during fermentation batches should be about minimum 5°C above coolant to maximum 80°C, display in 0.1°C increment with temperature rises of approx 1°C / minute.
- The vessel should get sterilized automatically on the bench without the need of external steam generator.

3. Sensors (Measurement & Control):

- All these parameters should have a separate probe which should be controlled by PID (Proportional-integral-derivative controller).
- **Temperature range** : 5°C to 80°C, displayed in 0.1°C increments using Platinum 1 RTD probe
- **pH range**: 2 - 12 pH via PID
- **pH Sensor**: Option for one or two gel pH probes with digital display in 0.01 increments
- **pH Control**: Cascade to pumps, gases and/or loops from external devices
- **DO range**: 0 - 200% via PID
- **DO Sensor**: Option of one or two polarographic DO probes with digital display in 0.1% increments
- **DO Control**: Cascade to agitation, gases, pumps and/or loops from external devices
- Foam/Level Sensor: 2
- Optional Sensors: Redox, DO

4. AGITATION:

- **Drive**: Top magnetic drive with single mechanical seal
- **Range & Control**: 50 - 1000 RPM, ±1 at 100 RPM; ±2 at 500 RPM; ±5 at 1000 RPM
- **Impellers**: Two six-bladed Rushton impellers
- **Baffles**: Four 316L removable stainless steel baffles

5. Ports:

For 7 lit

- **Headplate port**: 6.35 mm - 2; 12 mm - 9; 19 mm - 1; 13.5 mm - 2
- **Upper side wall**: 2" Tri-clamp
- **Bottom**: 0.75" NA connect

6. AERATION:

- **Gas System**: Standard: 1 TMFC with 0.5 to 25 SLPM (Standard litres per min) flow rate and built in four-gas control (Air, O₂, N₂ & CO₂). Optional: Rotameter or 2nd, 3rd or 4th TMFCs for individual gas control

- **Gas Inlet:** Ring sparger is provided with 0.2µl absolute disposable filter for use as a sparger or overlay
- 7. PUMPS:**
- **Standard:** Three built-in, assignable, peristaltic pumps and one external variable speed peristaltic pump for inoculums/feed
 - **Control modes:** Off, Prime, Base, Acid, Foam, Level 2 Wet, Level 2 dry.
 - **Speed:** Pumps 1 & 2: 10-12 RPM fixed speed, ability to view total pump flow rates. Pump 3: 100 RPM fixed speed, ability to view total pump flow rates.
 - **Inoculum/FEED (Variable Speed):** Watson marlow 0-200 RPM

8. PRESSURE:

- **Safety device:** A sanitary rupture disk in the vessel and an ASME safety release valve on the drain jacket

9. Sampling and Harvesting:

- Autoclavable sampling system with needle and luer lock system
- Harvest through bottom drain valve

10. Material:

- Vessel and Jacket Material should be ASME/CE certified SS 316L and internal material surface of the vessel is Ra value less than 0.55 micron outer surface to an Ra of less than 0.88 micron. Bioreactor should be SIP (Sterilization in place) design and all probes should be in-situ sterilizable.

11. Measurement and control:

- The system should have a powerful Reactor Process Controller providing integrated control of up to 32 process parameters; with a capability of trending 8 parameters simultaneously and storage of up to 10 recipes.
- The system should have large touchscreen interface that simplifies entering set points and provide easy-to-read displays of current values, cascade loops, sterilization cycles.
- The system should come with 2 USB ports for importing firmware / software upgrades and exporting trained data.

12. Peripherals and accessories:

- The system should have illuminated sight glass window of at least 1.5” and light for clear viewing of vessel contents.
- Sterilization through built-in band heater (no external steam generator required)
- Suitable Chiller should be provided for 7 L vessel or 1, chiller compatible and capable of running both 7L and 14 L vessel simultaneously should be provided
- Suitable Air Compressor compatible for 7L and 14 L vessel should be provided
- Online 5KVA UPS must be quoted which is capable of meeting the power demand of the entire system
- The above system should be cGMP compliant and validatable and should be capable of automatic sterilization using labs’ water supply (RO Water).

II Fermentor with 14L total capacity

1. Fermentor:

- In situ sterilizable fermentor with Working Volume 4 to 10 litre (Total Volume 14 litre respectively). Fermentor should be capable of accommodating 14 L and 19.5 L vessel
- System should be suitable for microbial fermentation in aerobic & anaerobic condition for batch, fed batch & continuous operation
- Aspect Ratio: 2:1
- The fermentor should have advanced synoptic touch screen interface providing easy to read displays of current values, cascade loops sterilization cycles, customize PI values for all process parameter etc.

2. Process Definition:

- Programmable in situ sterilization, the system should be supplied with Stainless Steel Exhaust Condensor on headplate.
- Fermentor should be capable of reaching sterilization temperature up to at least 121°C and temperature control during fermentation batches should be about minimum 5°C above coolant to maximum 80°C, display in 0.1°C increment with temperature rises of approx 1°C / minute.
- The vessel should get sterilized automatically on the bench without the need of external steam generator.

3. Sensors (Measurement and control):

- All these parameters should have a separate probe which should be controlled by PID (Proportional-integral-derivative controller).
- **Temperature range** : 5°C to 80°C, displayed in 0.1°C increments using Platinum 1 RTD probe
- **pH range**: 2 - 12 pH via PID
- **pH Sensor**: Option for one or two gel pH probes with digital display in 0.01 increments
- **pH Control**: Cascade to pumps, gases and/or loops from external devices
- **DO range**: 0 - 200% via PID
- **DO Sensor**: Option of one or two polarographic DO probes with digital display in 0.1% increments
- **DO Control**: Cascade to agitation, gases, pumps and/or loops from external devices
- **Foam/Level Sensor**: 2

4. AGITATION:

- **Drive**: Top magnetic drive with single mechanical seal
- **Range & Control**: 50 - 1000 RPM, ± 1 at 100 RPM; ± 2 at 500 RPM; ± 5 at 1000 RPM
- **Impellers**: Two six-bladed Rushton impellers
- **Baffles**: Four 316L removable stainless steel baffle

5. Ports:

For 14 lit

- **Headplate port**: 6.35 mm - 2; 12 mm - 10; 19 mm - 1; 13.5 mm - 2
- **Upper side wall**: 2" Tri-clamp
- **Bottom**: 0.75" NA connect

6. AERATION:

- **Gas System**: Standard: 1 TMFC with 0.5 to 25 SLPM (Standard litres per min) flow rate and built in four-gas control (Air, O₂, N₂ & CO₂). Optional: Rotameter or 2nd, 3rd or 4th TMFCs for individual gas control.
- **Gas Inlet**: Ring sparger is provided with 0.2µl absolute disposable filter for use as a sparger or overlay

7. PUMPS:

- **Standard**: Three built-in, assignable, peristaltic pumps and one external variable speed peristaltic pump for inoculums/feed
- **Control modes**: Off, Prime, Base, Acid, Foam, Level 2 Wet, Level 2 dry.
- **Speed**: Pumps 1 & 2: 10-12 RPM fixed speed, ability to view total pump flow rates. Pump 3: 100 RPM fixed speed, ability to view total pump flow rates.
- **Inoculum/FEED (Variable Speed)**: Watson marlow 0-200 RPM

8. PRESSURE:

- **Safety device**: A sanitary rupture disk in the vessel and an ASME safety release valve on the drain jacket

9. Sampling and Harvesting:

- Autoclavable sampling system with needle and luerlock system
- Harvest through bottom drain valve

10. Material:

- Vessel and Jacket Material should be ASME/CE certified SS 316L and internal material surface of the vessel is Ra value less than 0.55 micron outer surface to an Ra of less than 0.88 micron. Bioreactor should be SIP (Sterilization in place) design and all probes should be in-situ sterilizable.

11. Measurement and control:

- The system should have a powerful Reactor Process Controller providing integrated control of up to 32 process parameters; with a capability of trending 8 parameters simultaneously and storage of up to 10 recipes.
- The system should have large touchscreen interface that simplifies entering set points and provide easy-to-read displays of current values, cascade loops, sterilization cycles.
- The system should come with 2 USB ports for importing firmware / software upgrades and exporting trained data.

12. Peripherals and accessories:

- The system should have illuminated sight glass window of at least 1.5 inch and light for clear viewing of vessel contents.
- Sterilization through built-in band heater (no external steam generator required)

- Suitable Chiller should be provided for 14L vessel
- Suitable Air Compressor should be provided for 14 L vessel
- Online 5 KVA UPS system must be quoted which is capable of meeting the power demand of the entire system
- The above system should be cGMP compliant and validatable and should be capable of automatic sterilization using labs' water supply (RO Water).

1, Remote Control software controlling both the fermenters (7L and 14L) simultaneously

Remote Control Software:

- SCADA Software (Windows) for monitoring and control of entire process and parameters with security features capable of controlling both the fermenters (7L and 14L) simultaneously
- Desktop PC having Intel® Core™ i5-4570 with Intel HD Graphics 4600 (3.2 GHz, 6 MB cache, 4 cores, 4GB RAM, 1TB Harddisk, 18.5 inch monitor, 3-years warranty, labour and on-site service, for SCADA software
- Other analog devices to be integrated up to 7 number
- Software supplied should feature basic data management & monitoring capabilities. It should have ability to trend & control parameter set points, establish alarm setting and produce batch records and should be ideal for basic process management. It should also have added enhanced control features including a sophisticated programming module, custom synoptic display window and equipment lock-out feature

Essential Spares: (1 set for each fermentor)

- 8 Nos DO membrane with electrolyte,
- extra pH probe-2Nos
- silicon tubing 3.2mm- 40 Mtrs, 6.4mm- 40m Mtrs, 4.8mm-40mtrs,
- Extra Mechanical seal 2Nos complete with housing and stirring rod,
- septum-200Nos, extra Needle assembly-4Nos,
- Extra Air filter 10 Nos,
- Extra Exhaust Filter 10Nos,
- Orings and gasket -2 sets,
- probe storage rack-2Nos,
- Feed bottle with filters 500ml-8Nos, 1 Litre-4Nos

Optional items

- Optional Sensors: Redox, DO
- Single chiller for both 7 L and 14 L system as well as independent chillers for both system should be quoted. Also, both local and imported chillers should be quoted.

A. GENERAL REQUIREMENTS

1. Company should have dedicated application specialists with a proven track record to provide onsite training and should also have application-training center in India. Company must have to provide comprehensive training our technical staffs/students/Scientists. In case of breakdown of the system company must agree to provide their facilities for use.
2. Company should have a commitment to provide all required services for relocation and reinstallation of the system from RCB-Faridabad Campus to ATPC Faridabad Campus when the latter facility is ready.
3. Quoted model must have at least five installations in India and the bidders should submit a comprehensive list of installations of the same instrument with similar application. Bidder should also provide at least three user certificates from recognized research Institute/University in India where instrument has been extensively used for similar applications.
4. The equipment should have 60 months warranty from the date of handing over the fully functional unit to the Institute, against manufacturing defects of material and workmanship.
5. Price break up for warranty period over 12 months should be mentioned
6. The post- warranty CMC (after 5 years) should also be quoted for spares (including) and labor for the complete system, which includes all the accessories, supplied such as UPS, etc.
7. Quoted model must have at least five installation in India and vendor must provide the user list

Note:

- The price comparison shall be made taking into account the cost of the equipment and accessories with **5 years warranty and post warranty CMC for Five years**. Failure to comply this condition entails rejection of bids.
- The selected vendor should also agree to relocate and re-install the equipment at the ATPC Facility at Faridabad Biotech Cluster without any additional financial implication.
- All columns in Technical specifications sheet should be fully filled otherwise the tenders will be rejected.
- Relevant literature and publications from the original manufacturer that support the quoted model's ability to perform all of the above capabilities and specifications mentioned in the tender documents must be attached otherwise the tenders are liable to be rejected.
- UP TIME GUARANTEE: The firm should provide uptime guarantee of 95%.
- Downtime penalty Clause: During the Guarantee/warranty period, desired uptime of 95% of 365 days (24 h) is required. If downtime is more than 5%, the institute shall be entitled to impose penalty in the form of extended warranty period equal to twice the downtime period. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least FIVE YEARS after handing over the unit to the Centre. If accessories/other

attachments of the system are procured from the third party, then the vendor must produce cost of accessory/other attachment and the CAMC from the third party separately along with the main offer and the third party will have to sign the CAMC with the Institute if required.

- **SPARE PARTS:** The separate price list of all spares and accessories and consumables, if any, (Including minor) required for maintenance and repairs in future after guarantee/warranty period must be attached/enclosed along with the sealed quotation (Financial Bid) failing which quotation will not be considered.
- If any spares & accessories other than the price list attached/enclosed by the firm are required for future repair it will be borne by the firm only.

Technical Compliance sheet for Bioreactor (7L and 14 L total volume)
(Bidder should submit this filled document along with TECHNICAL BID only)

S.N o.	Fermentor Specification	Please mention if present (yes/no)	Technical details of the item	If 'yes' then mention reference page number on brochure
1	In situ sterilizable fermentor with Working Volume of 2 to 5L (7 L total volume). Fermentor should be capable of accommodating 14 L and 19.5 L vessel as well			
	System should be able to culture Bacteria, Yeast or Fungi.			
	Aspect Ratio: 2:1			
	The fermentor should have advanced synoptic touch screen interface providing easy to read displays of current values, cascade loops sterilization cycles, customize PI values for all process parameter etc.			
	The vessel should be interchangeable 7, 14 & 19.5 L stainless steel		Interchangeable vessels Capable of accommodating:,, sizes	
2	Process Definition			
	Programmable in situ sterilization, the system should be supplied with stainless steel exhaust condenser on headplate. Fermentor should be capable of reaching sterilization temperature up to at least 121°C and temperature control during fermentation batches should be about minimum 5°C above coolant to maximum 80°C, display in 0.1°C increment with temperature rises of approx 1°C / minute. The vessel should get sterilized automatically on the bench without the need of external steam generator.		Max. sterilization temperature:----- C Range of temp. control: ---- to----- C including increment value ---- C	
3	Measurement & Control			
	All these parameters should have a separate probe which should be controlled by PID			
	Temperature range : 5°C to 80°C, displayed in 0.1°C increments using Platinum 1 RTD probe		Temperature control range:---- C to ---- C with increment value-----	
	pH range : 2 - 12 pH via PID		pH range: ---- to---pH	
	pH Sensor : Option for one or two gel pH probes with digital display in 0.01 increments		pH sensor type:---- PID control: Yes/No No. of sensor-----	
	pH Control : Cascade to pumps, gases and/or loops from external devices			
	DO range : 0 - 200% via PID		DO range:---- to---- %	
	DO Sensor : Option of one or two Polarographic DO probes with digital display in 0.1% increments		DO sensor type PID control Yes/No	

	DO Control: Cascade to agitation, gases, pumps and/or loops from external devices			
4	AGITATION:			
	Drive: Top magnetic drive with single mechanical seal		Drive type and seal type: No. of seals:	
	Range & Control: 50 - 1000 RPM, ± 1 at 100 RPM; ± 2 at 500 RPM; ± 5 at 1000 RPM		RPM range: Control range:	
	Impellers: Two six-bladed rushton impellers		No. and type of impellers:	
	Baffles: Four 316L removable stainless steel baffles		No. and type of baffles:	
5	PORTS:			
	Headplate port: 6.35 mm - 2; 12 mm -9 ; 19 mm - 1; 13.5 mm - 2 (Head plate port: 15)		Size ----- No.----- Purpose-----	
	Upper side wall: 2" Tri-clamp		Size-----	
	Bottom: 0.75" NA connect		Size-----	
6	AERATION:			
	Gas System: Standard: 1 TMFC with 0.5 to 25 SLPM flow rate and built in four-gas control(Air, O ₂ , N ₂ & CO ₂). Optional: Rotameter or 2nd, 3rd or 4th TMFCs for individual gas control		TMFC with ----to----- SLPM flow rate	
	Gas Inlet: Ring sparger is provided with 0.2 μ l absolute disposable filter for use as a sparger or overlay			
7	PUMPS:			
	Standard: Three built-in, assignable peristaltic pumps and one external variable speed peristaltic pump for inoculums/Feed		No. of Pumps:-----	
	Control modes: Off, Prime, Base, Acid, Foam, Level 2 Wet, Level 2 dry.		No. of control modes:----- Type of control modes:-----	
	Speed: Pumps 1 & 2: 10-12 RPM fixed speed, ability to view total pump flow rates. Pump 3: 100 RPM fixed speed, ability to view total pump flow rates. Inoculum/FEED (Variable Speed): Watson marlow 0-200 RPM		Pump capacity: Pump flow rate: Pump capacity: Pump flow rate: Pump capacity: Pump flow rate:	
8	PRESSURE:			
	Safety device: A sanitary rupture disk in the vessel and an ASME safety release valve on the drain jacket are standard			
9	Sampling and Harvesting			
	Autoclavable sampling system with needle and luer lock		Sampling mode:-----	

	Harvest through bottom drain valve		Harvesting mode:-----	
10	Material			
	Vessel and Jacket Material should be ASME/CE certified SS 316L and internal surface of the vessel should be polished to a Ra Value less than 0.55 micron and outer surface to an Ra of less than 0.88 micron. Bioreactor should be SIP design and all probes should be of In-situ sterilization type.		Vessel and jacket material certification: ----- Internal surface Ra value:-- ---- Outer surface Ra value:---- --	
	Factory test report and calibration reports part of validation document		Yes/No	
11	Measurement and control			
	The system should have a powerful Reactor Process Controller providing integrated control of up to 32 process parameters; with a capability of trending 8 parameters simultaneously and storage of up to 10 recipes.		Process controller parameters:	
	The system should have large touchscreen interface: simplifies entering setpoints and provide easy-to-read displays of current values, cascade loops, sterilization cycles etc.		Touchscreen interface size:-----	
	The system should come with at least 2 USB ports for importing firmware / software upgrades and exporting data.		No. of USB ports:-----	
13	Peripherals and accessories			
	Suitable online 5KVA UPS (x2) or one combined UPS system must be quoted which is capable of meeting the power demand of the entire system (7 L and 14L fermentor)			
	The system should have illuminated sight glass window of at least 1.5 inches and light for clear viewing of vessel contents.			
	Sterilization through built-in band heater (no external steam generator required)			
	Suitable Chiller should be provided for both 7 L and 14L vessel		Temperature control mode and display type: Pump type and pressure: Pump flow rate: Working temp. range:----- to----- C	

			Temp.stability range: +/- -- -- C Setting resolution: Refrigerant used	
	Suitable Air Compressor should be provided for both 7L and 14 L vessel			
	The above system should be CGMP compliant and validatable and should be capable of automatic sterilization using labs' water supply (RO Water).			
12	1, Remote Control Software controlling both 7L and 14L fermentor			
	SCADA Software (Windows) for monitoring and control of entire process and parameters with security features		Capable of simultaneously running ----- number systems fully	
	Desktop PC having Intel® Core™ i5-4570 with Intel HD Graphics 4600 (3.2 GHz, 6 MB cache, 4 cores, 4GB RAM , 1TB Harddisk, 18.5 inch monitor,3-years warranty, labour and on-site service, for SCADA software		Bit computing ----- Processor ----- Storage -----TB RAM---- Multi SATA--- R/RW----DVD combo drive--- USB----Fire wire port---- LCD monitor---- with resolution----- Graphic card ---	
	Other analog devices to be integrated upto 7 number		Number of analog devices integrated ---- Details -----	
	Software supplied should feature basic data management & monitoring capabilities. It should have ability to trend & control parameter set points, establish alarm setting and produce batch records and should be ideal for basic process management. It should also have added enhanced control features including a sophisticated programming module, custom synoptic display window and equipment lock-out feature.			
13	Peripherals and accessories			
	Online 5KVA UPS system must be quoted which is capable of meeting the power demand of the entire system			
	The system should have illuminated sight glass window of at least 1.5 inches and light for clear viewing of vessel contents.			
	Sterilization through built-in band heater(no external steam generator required)			

	Suitable Chiller should be provided for both 7 Land 14L vessel		Temperature control mode and display type: Pump type and pressure: Pump flow rate: Working temp. range:----- to----- C Temp.stability range: +/- -- -- C Setting resolution: Refrigerant used	
	Suitable Air Compressor should be provided for both 7Land 14 L vessel			
	The above system should be CGMP compliant and validatable and should be capable of automatic sterilization using labs' water supply (RO Water).			
14	Optional			
	Redox sensors			
	Single suitable chiller to run both 7 Land 14 L fermentor simultaneously		Specification details	
	Single suitable Air Compressor for operatingboth 7Land 14 L fermentor		Specification details	

Price Compliance sheet for Bioreactor (7L and 14 L total volume)

(Bidder should submit this filled document along with *PRICE BID* only)

S.No	Fermentor Specification	Please mention if present (yes/no)	Technical details of the item	If 'yes' then mention reference page number on brochure	Price per unit
1	In situ sterilizable fermentor with Working Volume of 2 to 5L (7 L total volume). Fermentor should be capable of accommodating 14 L and 19.5 L vessel as well				
	System should be able to culture Bacteria, Yeast or Fungi.				
	Aspect Ratio: 2:1				
	The fermentor should have advanced synoptic touch screen interface providing easy to read displays of current values, cascade loops sterilization cycles, customize PI values for all process parameter etc.				
	The vessel should be interchangeable 7, 14 & 19.5 L stainless steel		Interchangeable vessels Capable of accommodating:,, sizes		
2	Process Definition				
	Programmable in situ sterilization, the system should be supplied with stainless steel exhaust condenser on headplate. Fermentor should be capable of reaching sterilization temperature up to at least 121°C and temperature control during fermentation batches should be about minimum 5°C above coolant to maximum 80°C, display in 0.1°C increment with temperature rises of approx 1°C / minute. The vessel should get sterilized automatically on the bench without the need of external steam generator.		Max. sterilization temperature: ---- C Range of temp. control: ---- to ---- C including increment value ---- C		
3	Measurement & Control				
	All these parameters should have a separate probe which should be controlled by PID				
	Temperature range : 5°C to 80°C, displayed in 0.1°C increments using Platinum 1 RTD probe		Temperature control range: ---- C to ---- C with increment value ----		
	pH range : 2 - 12 pH via PID		pH range: ---- to --- pH		
	pH Sensor : Option for one or two gel pH probes with digital display in 0.01 increments		pH sensor type: ---- PID control: Yes/No No. of sensor: ----		

	pH Control: Cascade to pumps, gases and/or loops from external devices				
	DO range: 0 - 200% via PID		DO range:---- to---- %		
	DO Sensor: Option of one or two Polarographic DO probes with digital display in 0.1% increments		DO sensor type PID control Yes/No		
	DO Control: Cascade to agitation, gases, pumps and/or loops from external devices				
4	AGITATION:				
	Drive: Top magnetic drive with single mechanical seal		Drive type and seal type: No. of seals:		
	Range & Control: 50 - 1000 RPM, ± 1 at 100 RPM; ± 2 at 500 RPM; ± 5 at 1000 RPM		RPM range: Control range:		
	Impellers: Two six-bladed rushton impellers		No. and type of impellers:		
	Baffles: Four 316L removable stainless steel baffles		No. and type of baffles:		
5	PORTS:				
	Headplate port: 6.35 mm - 2; 12 mm -9 ; 19 mm - 1; 13.5 mm - 2 (Head plate port: 15)		Size ----- No.----- Purpose-----		
	Upper side wall: 2" Tri-clamp		Size-----		
	Bottom: 0.75" NA connect		Size-----		
6	AERATION:				
	Gas System: Standard: 1 TMFC with 0.5 to 25 SLPM flow rate and built in four-gas control(Air, O ₂ , N ₂ & CO ₂). Optional: Rotameter or 2nd, 3rd or 4th TMFCs for individual gas control		TMFC with ----to-- ----SLPM flow rate		
	Gas Inlet: Ring sparger is provided with 0.2 μ l absolute disposable filter for use as a sparger or overlay				
7	PUMPS:				
	Standard: Three built-in, assignable peristaltic pumps and one external variable speed peristaltic pump for inoculums/Feed		No. of Pumps:----- -		
	Control modes: Off, Prime, Base, Acid, Foam, Level 2 Wet, Level 2 dry.		No. of control modes:----- Type of control modes-----		
	Speed: Pumps 1 & 2: 10-12 RPM fixed speed, ability to view total pump flow rates. Pump 3: 100 RPM fixed speed, ability to view total pump flow rates. Inoculum/FEED (Variable Speed): Watson marlow 0-200 RPM		Pump capacity: Pump flow rate: Pump capacity: Pump flow rate: Pump capacity: Pump flow rate:		
8	PRESSURE:				

	Safety device: A sanitary rupture disk in the vessel and an ASME safety release valve on the drain jacket are standard				
9	Sampling and Harvesting				
	Autoclavable sampling system with needle and luer lock		Sampling mode:---- --		
	Harvest through bottom drain valve		Harvesting mode:-- -----		
10	Material				
	Vessel and Jacket Material should be ASME/CE certified SS 316L and internal surface of the vessel should be polished to a Ra Value less than 0.55 micron and outer surface to an Ra of less than 0.88 micron. Bioreactor should be SIP design and all probes should be of In-situ sterilization type.		Vessel and jacket material certification: ----- Internal surface Ra value:----- Outer surface Ra value:-----		
	Factory test report and calibration reports part of validation document		Yes/No		
11	Measurement and control				
	The system should have a powerful Reactor Process Controller providing integrated control of up to 32 process parameters; with a capability of trending 8 parameters simultaneously and storage of up to 10 recipes.		Process controller parameters:		
	The system should have large touchscreen interface: simplifies entering setpoints and provide easy-to-read displays of current values, cascade loops, sterilization cycles etc.		Touchscreen interface size:-----		
	The system should come with at least 2 USB ports for importing firmware / software upgrades and exporting data.		No. of USB ports:-- -----		
13	Peripherals and accessories				
	Suitable online 5KVA UPS (x2) or one combined UPS system must be quoted which is capable of meeting the power demand of the entire system (7 L and 14L fermentor)				
	The system should have illuminated sight glass window of at least 1.5 inches and light for clear viewing of vessel contents.				

	Sterilization through built-in band heater(no external steam generator required)				
	Suitable Chiller should be provided for both 7 L and 14L vessel		Temperature control mode and display type: Pump type and pressure: Pump flow rate: Working temp. range:----- to----- C Temp.stability range: +/- ---- C Setting resolution: Refrigerant used		
	Suitable Air Compressor should be provided for both 7L and 14 L vessel				
	The above system should be CGMP compliant and validatable and should be capable of automatic sterilization using labs' water supply (RO Water).				
12	1, Remote Control Software controlling both 7L and 14L fermentor				
	SCADA Software (Windows) for monitoring and control of entire process and parameters with security features		Capable of simultaneously running ----- number systems fully		
	Desktop PC having Intel® Core™ i5-4570 with Intel HD Graphics 4600 (3.2 GHz, 6 MB cache, 4 cores, 4GB RAM , 1TB Harddisk, 18.5 inch monitor,3-years warranty, labour and on-site service, for SCADA software		Bit computing ---- Processor ---- Storage -----TB RAM---- Multi SATA--- R/RW---DVD combo drive--- USB---Fire wire port---- LCD monitor---- with resolution---- Graphic card ---		
	Other analog devices to be integrated upto 7 number		Number of analog devices integrated - --- Details -----		
	Software supplied should feature basic data management & monitoring capabilities. It should have ability to trend & control parameter set points, establish alarm setting and produce batch records and should be ideal for basic process management. It should also have added enhanced control features including a sophisticated programming module, custom synoptic display window and equipment lock-out feature.				

13	Peripherals and accessories				
	Online 5KVA UPS system must be quoted which is capable of meeting the power demand of the entire system				
	The system should have illuminated sight glass window of at least 1.5 inches and light for clear viewing of vessel contents.				
	Sterilization through built-in band heater(no external steam generator required)				
	Suitable Chiller should be provided for both 7 Land 14L vessel		Temperature control mode and display type: Pump type and pressure: Pump flow rate: Working temp. range:----- to----- C Temp.stability range: +/- ---- C Setting resolution: Refrigerant used		
	Suitable Air Compressor should be provided for both 7Land 14 L vessel				
	The above system should be CGMP compliant and validatable and should be capable of automatic sterilization using labs' water supply (RO Water).				
14	Optional				
	Redox sensors				
	Single suitable chiller to run both 7 Land 14 L fermentor simultaneously		Specification details		
	Single suitable Air Compressor for operatingboth 7Land 14 L fermentor		Specification details		

REGIONAL CENTRE FOR BIOTECHNOLOGY
NCR Biotech Science Cluster, Gurgaon-Faridabad Express Way,
Gurgaon-Faridabad Expressway 3rd Milestone towards Gurgaon,
Village-Bhankri, District Faridabad (Haryana) Tele No 0129-2876109

(SCHEDULE-'B')

Tender Ref. No. : RCB/ATN/10/14-15
Subject : Purchase of **BIO-REACTOR 7 & 14 ltrs.**
Date of Submission : On or before 21.04.2015 upto 14.00 hours.
Date of Opening : 21.04.2015 at 15.00 hours

1. Tender should be addressed to the Executive Director, Regional Centre for Biotechnology, Faridabad and submitted to the Office of the Regional Centre for Biotechnology under sealed cover failing which the tender shall be rejected. Terms and conditions for supply should invariably be indicated otherwise would be taken on its face value. The rates may be quoted on separate sheets failing which the tender(s) will be rejected.
2. Mixed quotations will not be considered for acceptance.
3. **IN CASE OF THE TENDER DOUMENTS DOWNLOADED FROM THE WEBSITE :-**
THE BIDDERS MAY DOWNLOAD THE TENDER DOCUMENTS DIRECTLY FROM THE WEBSITE AVAILABLE AT www.rcb.res.in. IN SUCH CASE, THE BIDDERS ARE REQUIRED TO SUBMIT THE TENDER COST FEE OF Rs. _____ (NON-REFUNDABLE) BY WAY OF SEPARATE DEMAND DRAFT DRAWN IN FAVOUR OF EXECUTIVE DIRECTOR, RCB, AND THE SAME SHOULD ESSENTIALLY BE ENCLOSED ALONGWITH THE TECHNO COMMERCIAL BID. THE BIDDERS SHOULD SPECIFICALLY SUPERSCRIBE, "DOWNLOADED FROM THE WEBSITE" ON THE TOP LEFT CORNER OF THE OUTER ENVELOPE CONTAINING TECHNO COMMERCIAL BID & PRICE BID SEPARATELY. IN NO CASE, THE TENDER COST FEE SHOULD BE MIXED WITH EMD AMOUNT. THE TENDERS NOT FOLLOWING THE ABOVE PROCEDURE WILL BE SUMMARILLY REJECTED.
4. TENDER SHOULD BE SUBMITTED IN TWO BID SYSTEM CONTAINING TWO PARTS AS DETAILED BELOW:
PART-I :- TECHNO-COMMERCIAL BID IN ONE SEALED COVER WITH E.M.D.
PART-II:- PRICE BID/FINANCIAL BID IN ONE SEALED COVER.

BOTH THE SEALED ENVELOPES SHOULD THEN BE PUT IN ONE OUTERCOVER INDICATING THEREON:

- i) Reference No. of the Tender _____
- ii) Tender regarding _____
- iii) Due date for submission of the tender : _____
- iv) Due date for opening of the tender _____
- v) Name of the firm _____

PLEASE NOTE THAT PRICES SHOULD NOT BE INDICATED IN THE TECHNO-COMMERCIAL BID. THE PRE-QUALIFICATION DOCUMENTS INCLUDING E.M.D./BID SECURITY AS REQUIRED IN THE TENDER DOCUMENT SHOULD INVARIABLY BE ACCOMPANIED WITH THE TECHNO-COMMERCIAL BID.

NOTE:-

- 1) TENDERS ARE TO DEPOSITED IN THE TENDER BOX KEPT AT THE RECEPTION OF RCB OFFICE, AFTER ENTERING THE DETAILS OF TENDER IN THE APPROPRIATE PAGE OF THE TENDER DEPOSIT REGISTER AVAILABLE WITH THE RECEPTIONIST. THOSE TENDERS WHICH ARE DIRECTLY PUT IN THE TENDER BOX WITHOUT DIARIZING IN THE TENDER DEPOSIT REGISTER ARE LIABLE TO BE REJECTED
- 2) TENDERS SUBMITTED WITHOUT FOLLOWING TWO BID SYSTEM PROCEDURE AS MENTIONED ABOVE WILL BE SUMMARILY REJECTED.

5. The tenderers should give rates, showing taxes, if any, and levies, packing forwarding and insurance charges separately giving full breakup details. THE INSTITUTE IS NOT AUTHORIZED TO ISSUE 'C/D FORMS'. PLEASE EXCLUDE CUSTOM DUTY/EXCISE DUTY COMPONENT IN RUPEE QUOTE AS THE INSTITUTE IS EXEMPTED FROM THE PAYMENT OF BASIC CUSTOMS DUTY/EXCISE DUTY. However, Excise Duty, if any, should be shown separately. Tender not confirming to this requirement shall be rejected and no correspondence will be entertained whatsoever.
6. IN CASE OF IMPORT THE TENDERERS ARE REQUIRED TO QUOTE FOB & CIF VALUE SEPARATELY DULY MENTIONING THE BREAK-UP DETAILS FOR FREIGHT & INSURANCE. THIS CONDITION SHOULD BE STRICTLY ADHERED TO, FAILING WHICH THEIR OFFER WILL BE SUMMARILY REJECTED.
The comparison between the indigenous and the foreign offers shall be made on FOR destination basis and CIF/CIP basis respectively. However, the CIF/CIP prices quoted by any foreign bidder shall be loaded further as under:
- a) Towards customs duty and other statutory levies –as per applicable rates with CDEC.
 - b) Towards custom clearance, inland transportation etc. - 2% of the CIF/CIP value.
7. **THE TENDERERS ARE REQUESTED TO SUBMIT THE FOLLOWING INFORMATION INVARIABLY TO MAKE PAYMENT THROUGH RTGS/NEFT. "The payment in pursuance of stores supplied/services rendered/work done will be made through RTGS/NEFT & charges incurred for affecting such electronic transfers will be borne by the vendors. The details of present charges for NEFT/RTGS are as under:**
- | | | |
|---------------------|----------------------|---------------------------|
| <u>NEFT:</u> | Up to Rs.1 lakh | – Rs.5/- per transaction |
| | Rs. 1 lakh and above | – Rs.25/- per transaction |
| <u>RTGS:</u> | Upto Rs.5 lakh | – Rs.25/- per transaction |
| | Rs.5 lakh and above | --Rs.50/- per transaction |
- To make payment through above said mode, the vendor/supplier/contracts have to submit the following information invariably:
- i. Name of the Beneficiary
 - ii. Account No. of the beneficiary
 - iii. IFCS Code of the Bank/Branch.
8. **THIS TENDER DOCUMENT IS NON-TRANSFERABLE.**
9. The bid document should be paged and a certificate may be provided on the covering letter indicating the number of pages submitted alongwith the bid.
10. The compliance report of specification should invariably be provided indicating the fulfillment of each parameter of the specifications failing which the offer will be rejected.
11. The checklist may be furnished properly and page No. may be mentioned against each Serial Number.
12. The Tender forms be clearly filled in ink legibly or type written giving full address of the tenderers. The tenderers should quote in figures as well as in words the rates amount tendered by him/them. Any discrepancy between the figures and words, the amount written in words will prevail. Alterations/over-writings, unless legibly attested by the tenderer, shall disqualify the tenders. The tenders should be signed by the tenderer himself/themselves or his/their authorized agent on his/their behalf (Authorization may be enclosed, if applicable).
13. THE FORWARDING LETTER/UNDERTAKING (SCHEDULE'A') DULY SIGNED SHOULD INVARIABLY BE ATTACHED ALONGWITH TECHNO-COMMERCIAL BID, FAILING WHICH THE TENDER SHALL BE REJECTED.
14. The tenderers should take care that the rates and amounts are written in such a way that interpolation is not possible, no blanks should be left which would otherwise, make the tender redundant.
- 15A. The tender rates should be kept open/valid for a period of one year from the date the tenders are opened.
15. The tenderers shall clarify/state whether he/they are manufacturer, accredited agent or sole representative indicating principals name & address. The offers of firms who are not manufacturer or direct authorized agent will be summarily rejected. **Sub-distributors will not be accepted.**

16. Delivery prospects with definite date of delivery at destination taking into cognizance transit facilities must be indicated.
17. EACH TENDER SHOULD BE ACCOMPANIED WITH AN EMD/BID SECURITY AMOUNTING TO **Rs. (as indicated in the tender Notice)**- BY WAY OF DEMAND DRAFT/ BANK GUARANTEE DRAWN IN FAVOUR OF “EXECUTIVE DIRECTOR, RCB”, (PREFERABLY BANK GUARANTEE) FAILING WHICH THE TENDER SHALL NOT BE CONSIDERED FOR ACCEPTANCE AND WILL BE OUTRIGHTLY REJECTED. IN CASE OF BANK GUARANTEE, IT SHALL BE VALID FOR ONE YEAR FROM THE DATE OF OPENING AND THE SAME SHOULD BE FROM ANY INDIAN NATIONALIZED BANK. **CASH/CHEQUE IS NOT ACCEPTABLE AT ALL.** THE EMD/BID SECURITY DEPOSITED AGAINST OTHER TENDERS CANNOT BE ADJUSTED OR CONSIDERED FOR THIS TENDER. NO INTEREST IS PAYABLE ON EMD/BID SECURITY.
18. **Liquidated Damages Clause :-**
- (a) The delivery date as stipulated shall be strictly adhered to failing which the Institute reserves the right to refuse the supplies. The extension of the date of delivery, if required should be obtained before the expected delivery date. The Institute also reserves the right to impose liquidated damages as enumerated below or to effect risk purchase on the firm’s cost and risk.
- (i) Supply delay for one week or a part thereof - @ 0.5%
- (ii) Supply delay for an additional week or a part thereof - @ 0.5% (subject to a ceiling of 10% of the order value.
- (b) In case of non-supply of items within a period of two months in excess of the stipulated delivery period, the order shall be cancelled and the following penalty shall be levied at the discretion of the Executive Director, RCB, GURGAON or his authorized representative:
- (i) Supply order of the value of Rs.100/- or below : No. Liq. Damages.
- (ii) Supply orders which remain unexecuted and total value of non-supplied items is above Rs.10000/-: @ 10% and/or administrative action, as deemed fit shall be taken against the defaulter.
19. **SOFT WARE AND HARDWARE UPGRADATION**
The selected firm for the supply of tendered item will have to provide free up-gradation of software (all update & upgrades) upto 5 years from the date of satisfactory installation.
20. Rates quoted should be valid for ONE YEAR from the date of opening of tender.
21. THE TENDERERS MAY BE REQUIRED TO DEMONSTRATE THE QUOTED MODEL OF THE EQUIPMENT DURING THE TECHNICAL EVALUATION, IF REQUIRED, FAILING WHICH THEIR BIDS/OFFER SHALL BE REJECTED. The firms are intimated that they should get ready for demonstration and only one-week time will be provided for arrangement of demonstration and no request for extending time for demonstration will be entertained. Failure to demonstrate, their offer will be summarily rejected.
22. Any other statutory levy imposed by the Govt. of India from time to time will be authorized on demand with adequate proof thereof will be paid extra.
23. Force majeure will be accepted on adequate proof thereof.
24. The Executive Director, RCB, Gurgaon shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions.
25. No payment shall be made for rejected Stores. Rejected items must be removed by the tenderers within two weeks of the date of rejection at their own cost and replace immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.
26. The Executive Director, RCB, reserves the right to cancel/reject full or any part of the tender which do not fulfill the conditions stipulated in the tender.
27. **Tenderers submitting tenders would be deemed to have considered and accepted all the terms and conditions. No enquiries, verbal or written, shall be entertained in respect of acceptance or rejection of the tender.**
28. TENDER MAY BE REJECTED IF THE COPY OF VALUE ADDED TAX (VAT) REGISTRATION IS NOT FURNISHED (IF APPLICABLE).
29. The quantity shown in the tender can be **increased or decreased** to any extent depending upon the actual requirement.
30. Any action on the part of the tenderer to influence anybody in the Centre, will be taken as an offence, he will not be allowed to participate in the tender enquiry and their offer will not be considered.

31. **Material confirming to the specifications should be quoted. Original Catalogue, Leaflets, literatures with full technical details should invariably be attached along with their offer, failing which their offer will be summarily rejected.**
32. **THE TENDERERS MUST QUOTE THEIR MODEL STRICTLY AS PER TENDER SPECIFICATIONS.**
33. Genuine equipments and instruments etc., should be supplied. Tenderers should indicate the source of supply i.e. name & address of the manufacturers from whom the items are to be imported, country of origin, country of Shipment etc.
34. The quoted equipment should preferably be of Indian/International standards.
35. The tenderers are required to quote the mode of shipment by Air/Sea/ Airport Parcel and should give separate breakup of freight and Insurance Charges.
36. Supply of equipment means-Supply, Installation and Commissioning at site. No separate charges will be paid separately on this account.
37. **UP TIME GUARANTEE:**
The firm should provide uptime guarantee of 95%.
38. **Downtime penalty Clause**
- 40.2 During the Guarantee/warranty period, desired uptime of 95% of 365 days (24 hrs) if downtime more than 5% the institute shall be entitled to impose penalty in the form of extended warranty period equal to twice the downtime period. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least **FIVE YEARS** after handing over the unit to the Centre. If accessories/other attachment of the system are procured from the third party, then the vendor must produce cost of accessory/other attachment and the CAMC from the third party separately along with the main offer and the third party will have to sign the CAMC with the Institute if required.
- 40.2 The principals or their agents are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.
- 41 **GUARANTEE/WARRANTEE PERIOD: THE TENDERERS MUST QUOTE FOR 5 YEARS COMPREHENSIVE WARRANTY (INCLUDING ALL SPARES, ACCESSORIES AND LABOR) FROM THE DATE OF COMPLETION OF THE SATISFACTORY INSTALLATION. THE WARRANTY CHARGES SHALL NOT BE QUOTED SEPARATELY OTHERWISE THE OFFER SHALL BE SUMMARITY REJECTED. ALSO THE BIDDERS ARE REQUESTED TO SUBMIT THEIR QUOTE (RATES) FOR SUBSEQUENT 5 YEARS COMPREHENSIVE AMC (INCLUDING ALL SPARES, ACCESSORIES AND LABOR). FAILURE TO COMPLY THIS CONDITION WILL ENTAIL THE REJECTION OF THE BIDS. THE PRICE COMPARISION SHALL BE MADE TAKING INTO ACCOUNT ON BASIC PRICE AND POST WARRANTY CMC.**
- 42 **Delivery:**The successful bidders should strictly adhere to the following delivery schedule supply, installation & Commissioning should be effected within 6 to 8 weeks from the date of supply order and this clause should be strictly adhere to failing which administrative action as deemed fit under rules will be taken against the defaulter.
- 43 **SPARE PARTS:** The separate price list of all spares and accessories and consumables, if any, (Including minor) required for maintenance and repairs in future after guarantee/warrantee period must be attached/enclosed along with the sealed quotation failing which quotation will not be considered.
If any spares & accessories other than the price list attached/enclosed by the firm are required for future repair it will be borne by the firm only.
- 44 The tenderers are required to furnish the list of spares along with their cost in the Financial bid failing which their bids are liable to be rejected.
- 45 The tenderer shall furnish a non-blacklisting certificate that the firm has not been blacklisted in the past by any government/Private institution. The tenderer/supplier has to give an affidavit on non-judicial stamp paper of Rs.10/- that there is no vigilance/CBI case pending against the firm/supplier and the firm has not been blacklisted in the past by any Govt. or Private Organization.
- 46 Payment of Agency Commission, if any, payable in rupees must be indicated. If no Agency Commission is admissible from the foreign suppliers must be indicated specifically.
- 47 Tenderers should clearly indicate the name of the Manufacturers/Beneficiary of the Letter of Credit, country of Origin, place of shipment/Airfreightment, etc.
- 48 Local agents quoting on behalf of their foreign suppliers must attach valid authority letter in their favour. In case of distributor, the firm should be direct distributor from the principal's. **Either the Foreign Principals/manufacturers**

or their authorized Indian Agent should participate in the tender but not both. Also one agent cannot represent two suppliers or quote on their behalf in the same tender. The sub-distributor authority by distributor will not be accepted at all. In case the firm is neither manufacturer nor direct authorized dealer of the manufacturer, such offers will be summarily rejected.

- 49 SUCCESSFUL TENDERERS WILL HAVE TO FURNISH PERFORMANCE BANK GUARANTEE FOR 10% CONTRACT VALUE FROM ANY INDIAN NATIONALIZED BANK VALID FOR THE WARRANTY PERIOD AND TWO MONTHS EXTRA alongwith the final proforma Invoice.
- 50 The rates quoted for the Stores/Equipments, under the reference, by the supplier shall in no event exceed the lowest price at which the suppliers of the Stores/Equipments of identical description are made to any other person/organization/Institution during the period and should attach an undertaking.

FALL CLAUSE

- 51 If, at any time, during the said period, the supplier reduce the said prices of such Stores/Equipment or sales such stores to any other person/organization/Institution at a price lower than the chargeable, he shall forthwith notify such reduction or sale to the Executive Director, Regional Centre for Biotechnology and the price payable for the Stores supplied after the date of coming into force of such reduction or sale shall stand correspondingly reduced.
- 52 Successful tenderers, should give pre-alert intimation prior to shipment notifying both the nominated clearing agents as well as the Institute.
- 53 The supplier shall furnish the following certificate to the Accounts Officer (Stores Accounts) alongwith each bill for payment for supplies made against in Rate Contract Tender.
- “I/We certify that the Stores of description identical to the Stores supplied to the government under the contract against Tender herein have not been offered/sold by me/us to any other person/organization/Institution upto date of bill/the date of completion of supplies against all supply orders placed during the currency of the tender/rate contract at the price lower than the institute under contract /against tender”.
- 54 The supplier shall furnish a list of organizations where the equipment, in question, has/have been supplied with the period during the last one year and performance certificate from such organization may also be provided.

55 A) PAYMENT TERMS (IN CASE OF IMPORTS)

- 1) Agency Commission, if any, will be paid to Indian agents in Indian Rupees, after satisfactory installation of equipment.
- 2) Letter of Credit will be opened on CIF value, as the case may be, for full amount after deducting the Agency commission/Technical Service Charges, if any.
- 3) **100% payment shall be released against presentation of shipping documents against a performance Bank Guarantee valid for a period of 62 months from the date of satisfactory Installation certificate issued by the user department for 10% value from any Indian Nationalized Bank along with final proforma Invoice. In absence of Performance Bank Guarantee, 90% payment will be released against presentation of shipping documents & balance 10% payment will be released after satisfactory Installation certificate issued by the user department and against submission of Performance Bank Guarantee of 10% order value valid for a period of 62 months from the dated of satisfactory Installation certificate issued by the user department.**

B) PAYMENT TERMS (IN CASE OF RUPEE OFFER)

100% PAYMENT WILL BE MADE AFTER SUPPLY, installation & satisfactory demonstration of the said equipment subject to submission of Performance Bank Guarantee for 10% order value valid for a period of 62months from the date of satisfactory Installation certificate issued by the user department, from any Indian Nationalized Bank. FAILING WHICH, 90% PAYMENT WILL BE RELEASED AFTER INSTALLATION & BALANCE 10% PAYMENT WILL BE RELEASED AFTER WARRANTY PERIOD.

- 56) **PRE-REQUISITES, IF ANY, FOR INSTALLATION OF THIS EQUIPMENT SHOULD BE CLEARLY INDICIATED IN THE TECHNICAL BID, FAILING WHICH IT WILL BE PRESUMED THAT NO SUCH PRE-REQUISTES ARE REQUIRED AND IT WILL BE THE COMPLETE RESPONSIBILITY OF THE VENDOR TO MAKE THE EQUIPMENT FUNCTIONAL WITHIN THE QUOTE PRICE.**